

# Verification of Compliance

Product Name : **Medical Type Switching Power Supply**  
Brand Name : **MEAN WELL**  
Model Number : **MSP-1000-X (X=12, 15, 24, 48)**  
Applicant : **MEAN WELL ENTERPRISES CO., LTD.**  
Address : **No.28, Wuquan 3rd Rd., Wugu Dist., New Taipei City 248, Taiwan (R.O.C.)**  
Report Number : **S3C-M020-1711-126**  
Issue Date : **September 4, 2023**  
Applicable Standards :

**Emission :**  
**EN 60601-1-2:2015/A1:2021:**  
**Home Healthcare Environment**  
**EN 55011:2016/A2:2021 Group 1, Class B**  
**EN IEC 61000-3-2:2019**  
**EN 61000-3-3:2013/A1:2019**

**Immunity :**  
**EN 60601-1-2:2015/A1:2021: Home Healthcare Environment**  
EN 61000-4-2:2009  
EN IEC 61000-4-3:2020  
EN 61000-4-4:2012  
EN 61000-4-5:2014/A1:2017  
EN 61000-4-6:2014  
EN 61000-4-8:2010  
EN IEC 61000-4-11:2020



Based on the MDR (Regulation (EU) 2017/745) and the specifications of the customer, one sample of the designated product has been tested in our laboratory and found to be in compliance with the EMC standards cited above.



## Central Research Technology Co.

11, Lane 41, Fushuen St., Jungshan Chiu,  
Taipei 104, Taiwan

Tel : 886-2-25872719

Fax : 886-2-25850364

(Jim Chien/Authorized Signatory)

Date: September 4, 2023

TAF 0905

FCC CAB Code TW1104, TW0019

NVLAP Lab Code 200575-0

ISED CAB Code TW0905

VCCI Accep. No. R-11527, C-11609, T-11441, G-10010, C-20010

T-20009, G-10614